

Certificate of Analysis

11 Triazines in Soil QC

Catalog Number: SCQC-1052
Lot Number: 200914
Manufacture Date: 09/14/2020

Expiration: 09/30/2024
Matrix: Soil
Hazards: Irritant

<u>Analyte</u>	<u>CAS</u>	<u>Analyte Purity</u>	<u>Gravimetric Concentration (mg/kg)</u>			<u>Acceptance Limits (mg/kg)</u>
Ametryne	834-12-8	99.5%	4.0	±	0.04	1.99 - 5.97
Atraton	1610-17-9	99.0%	3.9	±	0.04	1.96 - 5.88
Atrazine	1912-24-9	100%	4.1	±	0.04	2.03 - 6.09
Prometon	1610-18-0	98.3%	3.9	±	0.04	1.97 - 5.90
Prometryne	7287-19-6	99.5%	4.1	±	0.04	2.03 - 6.09
Propazine	139-40-2	98.7%	4.1	±	0.04	2.07 - 6.22
Secbumeton	26259-45-0	97.7%	4.1	±	0.04	2.03 - 6.10
Simetryn	1014-70-6	99.2%	4.2	±	0.04	2.08 - 6.25
Terbutylazine	5915-41-3	99.5%	4.1	±	0.04	2.03 - 6.09
Terbutryne	886-50-0	98.1%	3.9	±	0.04	1.96 - 5.89
Simazine	122-34-9	99.9%	0.67	±	0.01	0.333 - 1.00

This certified reference material (CRM) was manufactured and certified by NSI Lab Solutions according to quality procedures meeting our accreditation requirements of ISO 17034:2016 and ISO/IEC 17025:2017. Our certificates and scopes of accreditation may be viewed at www.anab.org.

Packaging, Storage, Instructions For Use

This CRM is packaged in a sealed container and must be stored at 2°C to 8°C. To use this CRM, allow it to reach room temperature.

This sample has been designed to be totally used. Do not subsample since intra-sample homogeneity cannot be assured. Transfer the entire contents of the sample vial and rinse the vial with 2 small aliquots of extraction solvent adding rinsates to the extraction vessel. Complete the analysis according to your normal procedures.

Report results in units of mg/kg based on a 30 g sample size. No dry weight correction is needed.

Certified concentration is based upon the gravimetric/volumetric true value when prepared according to instructions.

Acceptance limits are based upon USEPA solid and chemical Interlaboratory studies.

Traceability Information

Analyte Source Materials: The highest purity analyte source materials are used in the manufacture of this standard. The actual purity is referenced above.

Catalog Number: SCQC-1052

Lot Number: 200914

Method: This CRM was verified Volumetrically/Gravimetrically

Balance: All analytical balances are calibrated on a semiannual basis by an ISO 17025 accredited calibration laboratory and are traceable to NIST. Traceable Calibration Certificate available upon request.

All balances are checked daily by an in-house standard operating procedure. The weights used for this daily verification are calibrated annually by an ISO 17025 accredited calibration laboratory and are certified traceable to NIST. Certificate of Calibration and Traceability available upon request.

Thermometer: All thermometers are NIST traceable through thermometers that are calibrated annually by an ISO 17025 accredited calibration laboratory.

Glassware: All glassware used in the manufacture of our standards is Class A. An in-house standard operating procedure is used to verify all glassware prior to it being placed into service. Volumetric pipetors are calibrated every four months by an ISO 17025 accredited calibration laboratory.

Intended Uses

- Calibration of analytical instruments
- Validation of analytical methods
- Preparation of working level reference materials, i.e. "check standards"
- Detection limit studies

Ken Grzybowski

Ken Grzybowski, Organics Department Manager

Mark Hammersla

Mark Hammersla, President